# **Product** Data Sheet

# **Product Datasheet**

7OXV Oxygen

# **Document Purpose**

The purpose of this document is to present the performance specification of the 7OXV oxygen sensor.

This document should be used in conjunction with the Operating Principles (OP02) and the Product Safety Datasheet (PSDS 4).

The data provided in this document are valid at 20°C, 50% RH and 1013 mBar for 3 months from the date of sensor manufacture. Output signal can drift below the lower limit over time. For guidance on the safe use of the sensor and on sensor performance outside of these environmental limits, please refer to the Operating Principles.

Page 1 of 3

Doc. Ref.:7OXV.indd ECN I 4728 Issue 4 30th January 2017

-

# **Product** Data Sheet



## **Key Features and Benefits:**

- Robust, industry standard 7-Series packaging
- Compact size

## **Technical Specifications**

### **MEASUREMENT**

Measurement Range Maximum Overload 30% vol. O Output Signal\* Response Time (T90)\* Offset (3 minutes N<sub>2</sub>)\*

**Technology** | Electrochemical 0-25% vol. O<sub>2</sub> 0.195 - 0.25 mA in Air

<15 Seconds

<0.5% vol.  $O_2$  Can be considered linear in

many cases. See Operating Principles (OP-02) for further details.

#### **ELECTRICAL**

Recommended Load Resistor | 100 Ω

## **MECHANICAL**

Weight | <34 g Casing Material | ABS

Linearity

## **ENVIRONMENTAL**

Operating Temperature Range | -20°C to +50°C **Recommended Storage Temp** 0°C to 20°C Temperature Coefficient | 0.2% signal / °C Absolute Pressure Range | Atmospheric ± 10% **Differential Pressure Range** 40 mBar maximum Pressure Coefficient < 0.02% signal / mBar **Relative Humidity Range** 


Intermittent | 0 to 99%RH non-condensing Continuous | 15 to 99%RH non-condensing

### LIFETIME

Long Term Output Drift\* | <5% signal / year Expected Operating Life | 2 years in air

Storage Life 6 months in original packaging

# **Product Dimensions**



All dimensions in mm All tolerances ±0.15 mm unless otherwise stated

## **IMPORTANT NOTES**

Do not remove label. Do not solder to pins.

When installing the sensor into instrumentation, the sensor vent hole should not be blocked. The instrument should also be adequately vented.

If the sensor vent hole is blocked or if the instrument is not adequately vented, sensor performance will be compromised.

For further details, refer to Operating Principles OP02.

\* Specifications are valid at 20°C, 50% RH and 1013 mBar, using recommended circuitry. Performance characteristics outline the performance of sensors supplied within the first 3 months. Output signal can drift below the lower limit over time.

Doc. Ref.:70XV.indd ECN I 4728 Issue 4 30th January 2017

Page 2 of 3

# **Product** Data Sheet

#### **Poisons**

Sensors are designed for operation in a wide range of environments and harsh conditions. However it is important that exposure to high concentrations of solvent vapours is avoided, both during storage, fitting into instruments, and operation. When using sensors with printed circuit boards (PCBs), degreasing agents should be used before the sensor is fitted. Do not glue directly on or near the sensor as the solvent may cause crazing of the plastic.

## **Cross Sensitivity Data**

Toxic gases at TLV levels will have no cross-sensitivity effect on Oxygen sensors. At very high levels (i.e. percent levels), highly oxidising gases (e.g. ozone, chlorine) will interfere to the extent of their oxygen equivalent, but most other commonly occurring gases will have no effect.

#### Acid Gases

IMPORTANT NOTE: Acid gases such as  $CO_2$  and  $SO_2$  will be absorbed by the electrolyte and tend to increase the flux of oxygen to the electrode. This gives an enhanced oxygen signal of approximately 0.3% of signal per 1%  $CO_2$ . Oxygen sensors are not suitable for continuous operation in concentrations of  $CO_2$  above 25%.

### **SAFETY NOTE**

This sensor is designed to be used in safety critical applications. To ensure that the sensor and/or instrument in which it is used, are operating properly, it is a requirement that the function of the device is confirmed by exposure to target gas (bump check) before each use of the sensor and/or instrument. Failure to carry out such tests may jeopardise the safety of people and property.

Every effort has been made to ensure the accuracy of this document at the time of printing. In accordance with the company's policy of continued product improvement, the manufacturer reserves the right to make product changes without notice. The products are always subject to a programme of improvement and testing which may result in some changes in the characteristics quoted. As the products may be used by the client in circumstances beyond the knowledge and control of the manufacturer, we cannot give any warranty as to the relevance of these particulars to an application. It is the clients' responsibility to carry out the necessary tests to determine the usefulness of the products and to ensure their safety of operation in a particular application.

Doc. Ref.:7OXV.indd ECN I 4728 Issue 4 30th January 2017

