Hydrogen Chloride

HCI 3E 30

HCI 3E 30

FEATURES

Amperometric 3 electrode sensor cell Short warm up time Good zero stability Fast response 1:1 cross interference to HBr

TYPICAL APPLICATIONS

TLV-Monitoring Leak detection Pharmaceutical industry

PART NUMBER INFORMATION

MINI	1139-034-30009
4 series adaptation	1139-034-30049
7 series adaptation	1139-034-30079

The manufacturer deems the data contained herein as factual, and the opinions expressed are those of qualified experts based on the results of tests conducted. The above data can not be used as a warranty provision or representation for which the manufacturer assumes legal responsibility. The data are offered solely for consideration, investigation and verification. Any use of this information is subject to federal, state and local laws and regulations. Rev. 11/2011

HCI 3E 30


TECHNICAL SPECIFICATIONS

Measuring Range Sensitivity Range Zero Current at 20 °C Resolution at 20 °C Bias Potential Linearity	0–30 ppm 140 nA/ppm ± 60 nA/ ppm < ± 100 nA < 0.7 ppm 200 mV < 5% full scale	
Response Time at 20 ℃ t50 t90	< 30 s calculated from 4 min. exposure time < 70 s calculated from 4 min. exposure time	
Long Term Sensitivity Drift	< 3% per month	
Operation Conditions Temperature Range Humidity Range	-20°C to +40°C 15–95% r.H., non–condensing	
Effect of Humidity	high humidity causes HCI absorption	
Sensor Life Expectancy Warranty	> 24 months 12 months	

The manufacturer deems the data contained herein as factual, and the opinions expressed are those of qualified experts based on the results of tests conducted. The above data can not be used as a warranty provision or representation for which the manufacturer assumes legal responsibility. The data are offered solely for consideration, investigation and verification. Any use of this information is subject to federal, state and local laws and regulations.

HCI 3E 30


Temperature dependence of output signal:



The manufacturer deems the data contained herein as factual, and the opinions expressed are those of qualified experts based on the results of tests conducted. The above data can not be used as a warranty provision or representation for which the manufacturer assumes legal responsibility. The data are offered solely for consideration, investigation and verification. Any use of this information is subject to federal, state and local laws and regulations.

HCI 3E 30

Temperature dependence of zero reading:



The manufacturer deems the data contained herein as factual, and the opinions expressed are those of qualified experts based on the results of tests conducted. The above data can not be used as a warranty provision or representation for which the manufacturer assumes legal responsibility. The data are offered solely for consideration, investigation and verification. Any use of this information is subject to federal, state and local laws and regulations.

Rev. 11/2011

HCI 3E 30

CROSS SENSITIVITIES AT 20 °C

Gas	Concentration	Reading [ppm]
Alcohols	1000 ppm	0
Ammonia	100 ppm	0
Arsine	0.2 ppm	0.7
Carbon Dioxide	5000 ppm	0
Carbon Monoxide	100 ppm	0
Chlorine	5 ppm	<±0.1
Hydrocarbons	% range	0
Hydrogen	10000 ppm	0
Hydrogen Cyanide	20 ppm	7
Hydrogen Sulfide	20 ppm	60
Nitric Oxide	100 ppm	45
Nitrogen	100 %	0
Nitrogen Dioxide	10 ppm	<±0.5
Phosphine	0.1 ppm	0.3
Sulfur Dioxide	20 ppm	8

Notes:

1. Interference factors may differ from sensor to sensor and with life time. It is not adviseable to calibrate with interference gases.

2. This table does not claim to be complete. The sensor might also be sensitive to other gases.

The manufacturer deems the data contained herein as factual, and the opinions expressed are those of qualified experts based on the results of tests conducted. The above data can not be used as a warranty provision or representation for which the manufacturer assumes legal responsibility. The data are offered solely for consideration, investigation and verification. Any use of this information is subject to federal, state and local laws and regulations.

Rev. 11/2011

Safety Note

This sensor is designed to be used in safety critical applications. To ensure that the sensor and/or instrument in which it is used, are operating properly, it is a requirement that the function of the device is confirmed by exposure to target gas (bump check) before each use of the sensor and/or instrument. Failure to carry out such tests may jeopardize the safety of people and property.

Attention

Use of this range of sensors requires complete understanding of the instructions. Before using this range, please carefully read 'Application Notes'.

For further assistance on sensor selection and use, please contact a member of the Technical Sales team.

The manufacturer deems the data contained herein as factual, and the opinions expressed are those of qualified experts based on the results of tests conducted. The above data can not be used as a warranty provision or representation for which the manufacturer assumes legal responsibility. The data are offered solely for consideration, investigation and verification. Any use of this information is subject to federal, state and local laws and regulations.