

- ★ READ THIS INSTRUCTION MANUAL AND THE INSTRUCTIONS OF THE ASPIRATING PUMP PRIOR TO USING THIS PRODUCT.
- ★ DO NOT DISCARD CAREFULLY THIS INSTRUCTION MANUAL UNTIL ALL THE TUBES IN THIS BOX ARE USED UP.

1. PERFORMANCE:

Measuring Range	: 2-12 LB/MMCF
and Pump Strokes	: 2 pump strokes
Sampling Time	: 2 minutes
Colour Change	: Yellow → Yellowish Green
Detectable Limit	: 1 LB/MMCF
Operating Temperature	: 0-35 °C (32-95°F) (Temperature correction is necessary.)
Aspirating Pump	: Model AP-20, AP-20S, AP-20S, 400B, AP-1, AP-1S or 400A

⚠ CAUTION

1. THE DETECTOR TUBE CONTAINS CHEMICAL REAGENTS.
2. DO NOT TOUCH THESE REAGENTS DIRECTLY ONCE TUBES WERE BROKEN.
3. KEEP THE TUBES OUT OF THE REACH OF CHILDREN.

NOTICE

1. USE ONLY WITH PUMP MODELS AP-20, AP-20S, 400B, AP-1, AP-1S OR 400A. OTHERWISE, CONSIDERABLE ERROR IN INDICATION MAY OCCUR.
2. BEFORE TESTING, CHECK THE ASPIRATING PUMP FOR LEAKS. (REFER TO ITEM 7. INSPECTION OF ASPIRATING PUMP.) ANY PUMPS SHOWING SIGNS OF LEAKAGE SHOULD BE CORRECTED BEFORE USE.
3. DO NOT USE THIS TUBE OUTSIDE THE STATED OPERATING TEMPERATURE RANGE.
4. STORE TUBES IN A COOL AND DARK PLACE (0-25 °C/32-77°F), AND USE BEFORE EXPIRATION DATE PRINTED ON THE TOP OF THE BOX.
5. PRIOR TO USE, READ ITEM 8. USER RESPONSIBILITY CAREFULLY.
6. READ THE CONCENTRATION IMMEDIATELY AFTER DRAWING THE SAMPLE.

2. SAMPLING AND MEASUREMENT:




Fig.1

- ① Break both ends of the detector tube.

⚠ CAUTION SAFETY GLASSES AND GLOVES SHOULD BE WORN TO PREVENT INJURY FROM SPLINTERING GLASS.

- ② Insert the detector tube into the aspirating pump securely as shown in Fig.1. (Arrow mark shall point to the pump.)
- ③ Align the guide marks on the shaft and stopper of the aspirating pump.
- ④ Pull the pump handle at a full stroke until it locks and wait for 1 minute or until confirmation of sampling is completed with the flow indicator of the pump.(See descriptions about the flow indicator in the instruction manual of the pump).
- ⑤ Turn the pump handle right or left by 1/4 (90°), push back the handle without removing the detector tube from the pump. Then repeat the steps ③ to ④ once more.
- ⑥ On completion of sampling, read the scale at the top of the stained layer.

SPECIAL NOTE: I. The scale is calibrated at 20 °C (68°F), and 1013hPa. Readings obtained in other circumstances should be corrected. (REFER TO ITEM 3. CORRECTION FOR AMBIENT CONDITIONS.)

II. When the maximum point of the stained layer is unclear or oblique, read the scale at the centre between the longest and shortest points.

3. CORRECTION FOR AMBIENT CONDITIONS:

① Temperature; Readings obtained should be corrected with the following temperature correction table.

Table of the coefficient for temperature correction based on 20 °C)

Temperature(°C)	0	1	2	3	4	5	6	7	8	9
Correction Factor	1.69	1.65	1.61	1.57	1.53	1.50	1.46	1.42	1.39	1.35
Temperature(°C)	10	11	12	13	14	15	16	17	18	19
Correction Factor	1.32	1.28	1.25	1.22	1.19	1.15	1.12	1.09	1.06	1.03
Temperature(°C)	20	21	22	23	24	25	26	27	28	29
Correction Factor	1.00	0.98	0.95	0.92	0.90	0.87	0.85	0.82	0.80	0.78
Temperature(°C)	30	31	32	33	34	35				
Correction Factor	0.75	0.73	0.71	0.69	0.67	0.65				

Procedure of temperature correction: True concentration can be obtained by multiplying the readings of tubes by coefficient for temperature correction shown in the above. Therefore,

$$\text{True concentration (LB/MMCF)} = \text{Readings(LB/MMCF)} \times \text{Coefficient for temperature correction}$$


② Atmospheric Pressure ;

$$\text{True concentration} = \frac{\text{Temperature corrected} \times \text{concentration}}{\text{Atmospheric pressure (in hPa)}} \times 1013$$

4. INTERFERENCES:

Coexistence of Benzene, Toluene, Hexane, Sulphur dioxide and/or Hydrogen sulphide does not affect the readings. Alcohols produce a similar stain and give higher readings.

5. CHEMICAL REACTION IN THE DETECTOR TUBE:



6. DISPOSAL OF TUBES:

USED TUBES SHOULD BE DISCARDED CAREFULLY ACCORDING TO RELEVANT REGULATIONS, IF ANY.

7. INSPECTION OF ASPIRATING PUMP:

Checking for leaks;

- ① Insert a sealed, unbroken detector tube into the pump.
- ② Align the guide marks on the shaft and stopper of the pump.
- ③ Pull the handle to a full stroke and wait for 1 minute.
- ④ Unlock the handle and allow it to return slowly into the pump by holding the cylinder and handle securely.

⚠ CAUTION HANDLE WILL TEND TO SNAP BACK INTO THE PUMP QUICKLY.

- ⑤ If the handle returns completely to the original position, the performance is satisfactory. Otherwise, refer to maintenance procedures shown in the instruction manual of the pump to correct the leakage.

8. USER RESPONSIBILITY:

It is the sole responsibility of the user of this equipment to ensure that the equipment is operated, maintained, and repaired in strict accordance with these instructions and the instructions provided with each Model AP-20, AP-20S, 400B, AP-1, AP-1S or 400A aspirating pump, and that detector tubes are not used beyond their expiration date or have a colour change different to that stated in the Performance specifications.

The Manufacturer and Manufacturer's Distributor shall not be otherwise liable for any incorrect measurement or any damages, whether damages result from negligence or otherwise.

※ Product specifications are subject to change without any prior notice.

IME1774/3

Printed in Japan