

## 1. PERFORMANCE


- 1) Measuring range : 0.01-1.0 %
- Number of pump strokes : 2 (200mℓ)
- 2) Sampling time : 3 minutes/2 pump strokes
- 3) Detectable limit : 10 ppm
- 4) Shelf life : 3 years
- 5) Operating temperature : 0 ~ 40 °C
- 6) Reading : Graduations printed on the tube are calibrated by Methyl ethyl ketone at 2 pump strokes and Butyl acetate concentration is determined by using a conversion chart.
- 7) Colour change : Orange → Brownish green

## 2. RELATIVE STANDARD DEVIATION

RSD-low : 15 %    RSD-mid. : 10 %    RSD-high : 10 %

## 3. CHEMICAL REACTION

Dichromate is reduced.




## 4. CALIBRATION OF THE TUBE

GAS CHROMATOGRAPHY

## 5. INTERFERENCE AND CROSS SENSITIVITY

Substance	Interference	ppm	Coexistence
Acetylene		3 %	Whole reagent is changed to Brown.
Propane		0.2 %	∕
Other organic gases or vapours except Halogenated hydrocarbons	Similar stain is produced.	50	Higher readings are given.

Butyl acetate (%)



No.139SB Tube reading (%)